

तेजपुर विश्वविद्यालय / TEZPUR UNIVERSITY (संसद के अधिनियम द्वारा स्थापित केंद्रीय विश्वविद्यालय) (A Central University established by an Act of Parliament) कुल सचिव का कार्यालय/ OFFICE OF THE REGISTRAR नपाम :: तेजपुर - 784028 :: असम NAPAAM :: TEZPUR - 784028 :: ASSAM

CORRIGENDUM NOTICE (ET-NIQ- 3422 DT-29-11-2018)

The Technical Specification of the Notice Inviting Quotations (NIQ) No. ET-NIQ-3314-DT-20-11-2018 (Tender ID: 2018 TEZU 409652 1) has been revised.

(Revised Technical Specifications cum Compliance Report is attached/uploaded separately)

1

The other contents of the NIQ remain unchanged.

Upon Aprilie

Assistant Registrar (GA) Tezpur University

<u>Revised Technical Specifications cum Compliance Report</u> (To be submitted on Company's/Firm's Letterhead Signed and Sealed)

| SI.No. | Specification | ompany's/Firm's Letterhead Signo Range/ Details | Complied | Remarks, |
|--------|------------------------------|--|----------|----------|
| | | | (Yes/No) | if any; |
| 1 | - Fully computer controlled | | | |
| | steady state and time | | | |
| | resolved | | | |
| | photoluminescence set-up | | | |
| | | | | |
| | - The instrument should be | | | |
| | capable of measuring | | | |
| | fluorescence of liquid/solid | | | |
| | sample as a function of time | | | |
| | - The lifetime from | | | |
| | picosecond to nanoseconds | | | |
| | using time correlated single | | | |
| | photon counting (TCSPC) | | | |
| | system | | | |
| | | | | |
| | - Optional facility for low | | | |
| | temperature | | | |
| | measurements. | | | |
| 2 | Time range | 0.1ns to 100 μs (fluorescence | | |
| | | and phosphorescence) | | |
| 3 | Resolution | < 50 ps | | |
| 4 | Excitation Source | 1. Pulse width < 1ns | | |
| | LASER/ LED; | 2. UV (300-400 nm), visible | | |
| | | (500-700 nm) | | |
| 5 | Optics and | All refractive. | | |
| - | Spectral Range | 200 ~700 nm | | |
| 6 | Bandpass | ≤ 30nm; continuously | | |
| | Excitation Dance | adjustable | | |
| | Excitation Range | 200-950 nm optimized in UV with automatic filter for | | |
| | | second order removal | | |
| 7 | Wavelength accuracy | > ± 0.5 nm | | |
| 8 | Scan speed | 100 ~150 nm/s | | |
| 9 | Sensitivity (signal to Noise | > 10000:1 (FSD method) | | |
| - | Ratio) | () | | |
| 10 | Emission Detector | Spectral range: 300-1100 nm | | |
| | | with single photon counting | | |
| | | photomultiplier | | |
| 11 | Reference Detector | Calibrated Silicon photodiode | | |

| 12 | Software |
|----|----------------------------|
| | - Windows based |
| | Fluorescence software |
| | - simplified drop down |
| | menus for operation |
| | -Detector algebra to |
| | customize data acquisition |
| | -Polarization control, |
| | automatic calculation for |
| | anisotropy study |
| | -Chromaticity/color |
| | coordinates calculation |